



ANAL FISTULA WO

Magnet Vendor/Field: SIEMENS 3.0T

Approved By: Dr. Quiroz

Published/Rev: 6/2026 / Rev 1

Seq	Pulse	Dim	Ori	Scan Dir	Phase Enc	FOV (mm)	Slc (mm)	Gap (mm)	TR	TE	Fat Sat	Freq	Phase	Notes
localizer--	fl	2D	Sag	R>>L	A>>P	420	6.0	1.2	7.8	3.69	None	256	205	
CORONAL T2 HASTE	haste	2D	Cor	A>>P	R>>L	360	6.0	1.2	2000	81.0	None	320	256	
SAGITTAL T2 NO FS	tse	2D	Sag	R>>L	A>>P	190	3.0	0.0	4040	94.0	None	256	205	
CORONAL T2 FS	tse	2D	Cor	A>>P	R>>L	190	3.0	0.0	5610	120.0	FS	320	320	
AXIAL T2 NO FS	tseR	2D	Axial	S>>I	R>>L	180	3.0	0.0	7180	98.0	None	320	256	
AXIAL T2 FS	tse	2D	Axial	S>>I	R>>L	180	3.0	0.0	7610	98.0	FS	320	256	