



AXILLA WO

Magnet Vendor/Field: SIEMENS 3.0T

Approved By: Dr. Quiroz

Published/Rev: 6/2026 / Rev 1

Seq	Pulse	Dim	Ori	Scan Dir	Phase Enc	FOV (mm)	Slc (mm)	Gap (mm)	TR	TE	Fat Sat	Freq	Phase	Notes
localizer	fl	2D	Axial	S>>I	A>>P	450	6.0	6.0	7.8	3.69	None	256	230	
SAG & COR LOC	fl	2D	Sag	R>>L	A>>P	300	6.0	3.0	7.8	3.69	None	256	192	
AXIAL T2 STIR	tir	2D	Axial	S>>I	R>>L	320	4.5	1.35	4090	99.0	IR	256	256	TI 215ms
SAGITTAL T2_tse_fast_dixon	tse	2D	Sag	Lat>>Med	H>>F	280	4.5	1.35	3000	66.0	Dixon	320	272	
CORONAL T1	tse	2D	Cor	A>>P	F>>H	280	5.0	1.5	500	9.4	None	320	240	
CORONAL T2 STIR BLADE	tir	2D	Cor	A>>P	F>>H	300	5.0	1.5	6080	78.0	IR	256	256	TI 210ms