

<b>INDICATION</b>		SOB, R/O mass, lung nodule						
<b>POSITION / LANDMARK</b>		Supine / Chin						
<b>START/END LOCATIONS</b>		Above apex through lungs						
<b>CONTRAST PARAMETERS</b>		na						
<b>RESPIRATORY PHASE</b>		Inspiration						
<b>SCAN DELAY</b>		na						
<b>SCAN TYPE</b>		Helical						
	<b>KV</b>	<b>mA</b>	<b>Rot Time (sec)</b>	<b>Pitch</b>	<b>Speed (mm/rot)</b>	<b>Noise Index</b>	<b>ASiR</b>	<b>Dose Reduction</b>
	120	Smart mA 50-500	0.5	0.984:1	39.37	27		
<b>TECHNIQUE</b>		Using automated exposure control and adjustment of the mA and/or kV according to patient size, radiation dose to be kept as low as reasonably achievable to obtain optimal diagnostic quality images.						
<b>Scans</b>								
<b>Series #</b>	<b>Series</b>	<b>Body Part</b>	<b>DFOV</b>	<b>Thick/Space</b>	<b>Algorithm</b>	<b>Notes</b>		
1	Loc					AP/Lat		
2	Source data wo	Chest	See note	5x5	STND	FOV open to cover bilateral axilla for all series		
<b>Recons</b>								
<b>Recon source Series #</b>	<b>Recon</b>	<b>Body Part</b>	<b>Thick / Space</b>	<b>Algorithm</b>	<b>W/L</b>	<b>Notes</b>		
2	AX	Chest	5x5	Mediastinum				
2	AX	Chest	2.5x2.5	Lung	2000/ -400			
2	SAG	Chest	3x3	Mediastinum				

\*Please note, recons are displayed as thickness X spacing



# CT CHEST without GE Revolution

Approval: M. Gange, DO

rev:2 2/12/2026

2	COR	Chest	3x3	Mediastinum		
2	AX	Chest	10x2	Lung	2000/ -400	
<b>2D / 3D Processing</b>						
<b>Series required in PACS</b>						
Loc, Dose Report, Source data, ALL recons						

**ADDITIONAL INSTRUCTIONS:**

\*Please note, recons are displayed as thickness X spacing