

<b>INDICATION</b>		Liver specific: Abnormal LFT's, Liver mass, Liver hemangioma, Triple phase Liver, Cirrhosis.						
<b>POSITION / LANDMARK</b>		Supine / Xiphoid						
<b>START/END LOCATIONS</b>		Above diaphragm through liver						
<b>CONTRAST PARAMETERS</b>		100cc Isovue 300, 2cc per second						
<b>RESPIRATORY PHASE</b>		Inspiration						
<b>SCAN DELAY</b>		30 sec, 60 sec, 3 minute						
<b>SCAN TYPE</b>		Helical						
<b>Series</b>	<b>KV</b>	<b>mA</b>	<b>Rot Time (sec)</b>	<b>Pitch</b>	<b>Speed (mm/rot)</b>	<b>Noise Index</b>	<b>ASiR</b>	<b>Dose Reduction</b>
with	120	100-440	0.7	0.984:1	39.37	19.63		
<b>TECHNIQUE</b>		Using automated exposure control and adjustment of the mA and/or kV according to patient size, radiation dose to be kept as low as reasonably achievable to obtain optimal diagnostic quality images.						
<b>Scans</b>								
<b>Series #</b>	<b>Series</b>	<b>Body Part</b>	<b>DFOV</b>	<b>Thick/Space</b>	<b>Algorithm</b>	<b>Notes</b>		
1	Loc					AP/Lat		
2	30 Sec	Abd	50	3.75 x 3.75	STND	FOV to include abd soft tissue		
3	60 Sec	Abd	50	3.75 x 3.75	STND	FOV to include abd soft tissue		
4	3 Minute	Abd	50	3.75 x 3.75	STND	FOV to include abd soft tissue		
<b>Recons</b>								
<b>Recon source Series #</b>	<b>Recon</b>	<b>Body Part</b>	<b>Thick / Space</b>	<b>Algorithm</b>	<b>W/L</b>	<b>Notes</b>		
3	AX	Abd	1.25x0.625	STND		Do not send to PACS, use for recons		
<b>AX 1.25x0.625</b>	COR	Abd	4 x 4	STND	400/40			

\*Please note, recons are displayed as thickness X spacing

